

## PRCSMADAD 01 – Clarifications

- The deadline for submission of tenders is 12:00 Monday 25 June.
- It is acceptable to submit a single tender guarantee as long as it covers the total amount requested of all the lots for which the respective tenderer is submitting bids.
- It is acceptable to submit the tender guarantee using the template found in the Practical guide of the European Commission, found here: <http://ec.europa.eu/europeaid/prag/document.do?nodeNumber=4.7> . It is also acceptable to submit the tender guarantee in an alternative format, as long as the guarantee within is of equal effect and function. The tender guarantee form has been uploaded on Daleel Madani.
- Tenderers shall not modify and/or submit the Draft contract, the Administrative compliance grid, the Evaluation grid and the certificates for provisional/final acceptance. They are included in the tender dossier as additional information for potential tenderers.
- No general clarification of the rules of origin and nationality will be provided. Because a specific question related to EU member states was received; EU member states are all eligible countries of origin.
- In the Contract notice, under the Selection and Award criteria on page 3 the following formulation can be found: “minimum annual turnover requested may not exceed 2 times the estimated annual contract value”. Please disregard this sentence, it is directed to the entity that publishes the call for tenders.
- As per the Contract notice tenders must remain valid for 90 days after the deadline for submission of tenders.
- Payments will be made as specified in the Special conditions. 40% per contract will be paid in advance, 60% per contract will be paid upon preliminary acceptance of all tasks.

## Related to technical specifications

- The Steam sterilizer must be FDA **or** CE approved.
- For the steam sterilizer PLC is **not** accepted.
- Anesthesia machine: ”Control For Anesthetic Gases (Automatic Change)” refers to the provision of safe monitoring of anesthetic gases O<sub>2</sub>; N<sub>2</sub>O and anesthetic agent.
- “Operating Ceiling lamp LED double head” – the heads are required to have two and three light bulbs respectively. This information is readily available in the table at the end of the Technical specification document.
- All specifications are based on the professional preferences of our medical and technical staff. We received an inquiry about the formulation of the specifications for the “Operating Ceiling lamp LED double head”. We would like to clearly state that the shape of the light requested is available from several suppliers, and we do not accept the claim that the specifications are shaped to suit the interests of a certain supplier. The choice to request boards with fewer light bulbs and with the shape mentioned has been made primarily from a service cost-efficiency perspective.